



Corrigendum

Corrigendum to: “Paper microfluidic device for early diagnosis and prognosis of acute myocardial infarction via quantitative multiplex cardiac biomarker detection” [Biosensors and Bioelectronics 128 (2019) 176–185]



Wei Yin Lim^a, T. Malathi Thevarajah^b, Boon Tong Goh^c, Sook Mei Khor^{a,d}

^a Department of Chemistry, Faculty of Science, University of Malaya, 50603 Kuala Lumpur, Malaysia

^b Department of Pathology, Faculty of Medicine, University of Malaya, 50603 Kuala Lumpur, Malaysia

^c Low Dimensional Materials Research Centre, Department of Physics, Faculty of Science, University of Malaya, 50603 Kuala Lumpur, Malaysia

^d University Malaya Centre for Ionic Liquids (UMCIL), University of Malaya, 50603 Kuala Lumpur, Malaysia

The authors regret to inform that the phrase used “99th percentile” should be omitted from the abstract as the intra- and inter-day precision study for cTnT determination was conducted at 10 ng mL⁻¹ instead of

at the clinical cutoff value of 0.01 ng mL⁻¹.

The authors would like to apologise for any inconvenience caused.

DOI of original article: <https://doi.org/10.1016/j.bios.2018.12.049>

E-mail address: naomikhor@um.edu.my (S.M. Khor).

<https://doi.org/10.1016/j.bios.2019.01.036>