



Influence of flexural rigidity on micromotion at the head-stem taper interface of modular hip prostheses[☆]



Henning Haschke*, Tobias Konow, Gerd Huber, Michael M. Morlock

Institute of Biomechanics, TUHH Hamburg University of Technology, Denickestraße 15, 21073 Hamburg Germany

ARTICLE INFO

Article history:

Received 21 November 2018

Revised 25 March 2019

Accepted 31 March 2019

Keywords:

Flexural rigidity

Micromotion

Relative motion

Taper interface

Taper size

Modular hip prostheses

ABSTRACT

Fretting corrosion as one reason for failure of modular hip prostheses has been associated with micromotion at the head taper junction. Historically the taper diameter was reduced to improve the range of motion of the hip joint. In combination with other developments, this was accompanied by increased observations of taper fretting, possibly due to the reduced flexural rigidity of smaller tapers. The aim of the study was to investigate how the flexural rigidity of tapers influences the amount of micromotion at the head taper junction.

Three different stem and two different taper designs were manufactured. Experimental testing was performed using three different activity levels with peak loads representing walking, stair climbing and stumbling. The relative motion at the head-stem taper was measured in six degrees of freedom. Micromotion was obtained by subtraction of the elastic deformation derived from monoblock and finite element analysis.

Less rigid tapers lead to increased micromotion between the head and stem, enlarging the risk of fretting corrosion. The influence of the stem design on micromotion is secondary to taper design. Manufacturers should consider stiffer taper designs to reduce micromotion within the head taper junction.

© 2019 IPEM. Published by Elsevier Ltd. All rights reserved.

1. Introduction

The introduction of modular femoral head-stem combinations for total hip arthroplasties (THAs) in the 1980s established several advantages over monoblock designs. A combination of different materials for the head and the stem could be realized, which lead to a variety of advantageous combinations, such as cobalt chrome or ceramic heads on titanium stems. With a broader choice of materials, the head-stem combination could be matched to fit the patients' requirements, such as avoidance of allergic reactions to cobalt chrome or consideration of a high body mass index (BMI). Moreover, the opportunity to take benefit from the adaptation to the patient-specific anatomy concerning the centrum collum-diaphysis angle (CCD), the range of motion (ROM) and femoral offset was also advantageous [1–3].

This modularity was enabled by the introduction of conical modular press fit taper junctions. The connection between the femoral stem and the head is realized by a conical press-fit junction. The characteristics of the taper junction vary among different manufacturers and even for products of the same manufacturer as

no standardization of the taper junction is available [4,5]. In general, taper designs (e.g. 10/12, 11/13, 12/14 or 14/16) are labeled according to a theoretical ratio between diameters of two taper cross sections at a distance of 20 mm [6]. For 12/14 taper this would result in a theoretical taper angle of 5° 43' 30'', while the V40 design has a desired taper angle of 5° 40'. In this context it has to be noted, that due to individual manufacturing tolerances the female and male tapers angles are different to guarantee a defined but manufacturer specific clamping at either side of the taper. The variety of taper designs includes different surface structures, taper lengths and taper angles which can be combined with heads of different sizes and offsets [7], yet evidence-based recommendations regarding optimal taper connection design remain inconclusive and conflicting [8].

Innovations were not just introduced for head material and taper design, but also included a large variety of stem designs and materials. More than 100 different stem designs are commercially available [9,10]. Between the different designs, large variations of the length, diameter, shape, surface structure, coating or stem modularity can be found [11]. Typically, proximal stem designs can be separated into three prominent groups: round, oval and rounded square neck geometry (Fig. 1). The head-neck ratio has a major influence on the impingement behavior [2] and consequently defines the anatomical ROM of the affected joints. Previous research showed that a rounded square/trapezoidal neck

[☆] Hamburg, November 12th, 2018 (revised March 25th, 2019)

* Corresponding author.

E-mail address: haschke@tuhh.de (H. Haschke).




Fig. 1. Examples for the geometrical variations in the neck design of modular hip prostheses. The cross section varies continuously with distance from the proximal taper surface.

design increases the joint mobility in flexion and extension by 20% compared to round neck designs [12]. Furthermore, the combination of small cross-sectioned tapers or necks with large diameter heads increases the ROM without the risk of an impingement [13,14].

As many technical innovations come with a drawback, a new phenomenon was described in the 1990s known as metallosis [15–17]. Several reports of patients with high metal ions blood levels and metal ion accumulation in the surrounding tissue identified the modular taper junction as the root of cause [18–20]. In vivo, the taper junctions are exposed to an exigent environment of the human body such as tissues and body fluids. The national joint replacement registry of the Australian Orthopaedic Association (AOA) indicated that 25.6% of all revision surgeries with primary diagnosis osteoarthritis had to be performed due to periprosthetic aseptic loosening [21]. The bulk materials used are themselves highly biocompatible. However, released nano- and microparticles of these materials result in a negative response of the immune system [18]. The result is an increased activation of osteoclasts that can lead to a resorption of the surrounding bone material. This causes a loosening of the femoral stem or the formation of pseudo tumors and necessitates an early revision surgery of the prosthesis [22].

The loosening of the artificial joint is a result of tribological strains promoted by high cyclic loading, e.g. walking activities of 1–2 million steps per year [23,24]. During activities of daily living the artificial joint is loaded with high forces exceeding a

multiple of the body weight [25]. Induced by these loadings, micromotion between the taper surfaces of the femoral head and the femoral stem may occur. Micromotion at the head-stem taper junction in combination with the surrounding body fluid may lead to an abrasive release of the corrosive resistant passivation layer (Fig. 2) which is known as fretting corrosion [26]. Metal debris resulting from the fretting process, was linked to adverse local tissue reactions, systematic effects on the body and subsequent failure of the implant [18,20].

The drive to eliminate limitations in joint movement led to the introduction of thinned and shortened taper and neck geometries over several decades from the 1980s to 2010s. Porter et al. [10] found a wide variability in flexural rigidity of various taper designs and described a trend towards shorter and less rigid taper junctions with time. Regardless of the trend, several publications indicate an increased risk for fretting corrosion for less rigid and smaller tapers as well as stems [27–29]. Flexural rigidity as a prevalent risk factor for corrosion is mainly present due to micromotion that is promoted by higher elastic strains [26,27,30]. In a finite element study, Dyrkacz and colleagues showed significant differences in the amount of micromotion between taper sizes [31] and the increase in fretting corrosion might even be larger if materials with a lower modulus of elasticity are used for the stems, e.g. Ti6Al4V or TMZF vs. CoCr29Mo [27]. The reduced flexural rigidity and usage of titanium for femoral stems has been identified as a potential cause for the increased incidence of corrosion in current THA designs over a wide range of retrieval studies [10,32,33], while the effect of smaller taper sizes as a risk factor for early revision is discussed controversially [29,34]. As a change of geometry does not only affect the bending and torsion rigidity properties, but also the ROM, the reduction of the taper size to obtain an increased ROM, results in a trade-off in rigidity. A balance between flexural rigidity of the junction and impingement-free ROM must be achieved [35]. However, the influence of reduced flexural rigidity and elastic modulus of the stem on corrosion risk may also depend on other factors, e.g. the head material [32].

The variety of taper sizes in clinical use is still large. The range covers taper sizes of 8/10mm up to 14/16mm. The Exeter V40 (Stryker) and the Corail (Depuy Synthes) are the most frequently used femoral stems for total conventional hip replacement in Australia (AOA, 2017). The V40 taper shows significantly smaller dimensions than the 12/14 taper of the Corail stem. Besides the taper geometry, large variations can be seen in the design of the neck. Either of these two parameters can have an impact on micromotion and fretting corrosion and is therefore of great clinical relevance. The purpose of this study was to clarify whether lower flexural rigidity of the taper and the neck geometry increases the amount of micromotion at the head-stem taper junction under physiological loading conditions.




Fig. 2. Formation of fretting corrosion and the release of micro particles within the modular taper junction.




Fig. 3. Overview of custom-made proximal stem and stem taper designs. The transversal bore was added for handling purposes during disassembly. The round proximal stem design was used to investigate the influence of taper design (12/14, V40, long 12/14). The 12/14 taper designs were used to investigate the influence of proximal stem geometry (oval, round, rounded square).

2. Materials and methods

2.1. Materials and surface geometry

Based on the design of two clinically successful stem designs (Accolade TMZF, Stryker, Kalamazoo, MI, US; Corail, DePuy Synthes, Raynham, MA, US), five different models of the stem taper and the proximal stem section were manufactured from Ti6Al4V (Fig. 3; $n=3$ each). Geometries of the Accolade and Corail stems were captured via 3D surface scanning technique (Handyscan 700, Creafom, Lévis, Canada; 480,000 data points/s, measurement temperature 21 °C) with a resolution of 0.05 mm and an accuracy of 0.03 mm. This accuracy resulted in an error of less than 1/1000 of the neck geometry (Fig. 4). The 3D surface scans were further edited to remove minor surface irregularities which occurred during the scanning process (spatial filter; Geomagic Studio, Geomagic, Morrisville, NC, US). The male stem taper geometries of the Accolade and Corail were additionally acquired using a tactile coordinate measuring machine to increase accuracy (1.7 μm ; Crysta Apex S 574, Mitutoyo Germany, Neuss, Germany).

The influence of three different taper designs was determined based on a round proximal stem geometry ($\varnothing=14$ mm): a 12/14 taper, a V40 taper and a long 12/14 taper (Fig. 3, Table 1). Three different proximal stem designs consisting of the Accolade (oval), the Corail (rounded square) and a custom design with a round proximal geometry, all with a 12/14 taper, were used to investigate

the influence of proximal stem design (Fig. 3). The use of custom-manufactured models rather than original prostheses was chosen to facilitate the systematic analysis of the influence of geometric differences in the taper and proximal stems regions while neglecting factors such as elastic modulus, taper surface structure or taper angle.

2.2. Computation of flexural rigidity

Flexural rigidity is commonly determined based on either the diameter of the distal taper contact [27–29,36,37] or the geometric taper centroid diameter [38] for a specific position and a fixed diameter. In the present study, a continuous calculation of rigidity was performed along the taper axis for the taper and neck region. Hence, the calculation of the flexural rigidity is carried out in lateral and central direction at every single position of the whole geometry of the prosthesis (Fig. 5). To achieve this, the CAD models were converted into surface models and post-processed with Matlab (2017b, The Mathworks, Natick, MA, US). Surface geometries were then converted into a point cloud. Starting at the proximal taper plane, the local second moment of inertia was determined for any cross-section with an increment of 0.5 mm moving in distal direction along the taper axis. The local second moment of inertia (I) was then multiplied by the elastic modulus (E) to report the flexural rigidity (FR) of the proximal stem region. The elastic modulus used for the models was 120 GPa. To investigate the influence of material, a FE model was also created for the oval V-40 model using the material properties of TMZF (elastic modulus: 85 GPa), the material from which the Accolade stem is made.

2.3. Relative motion measurement

Prior to testing, each stem was cleaned and assembled quasi-statically ($F=2000$ N) with a metal ball head along the taper axis (ISO 7206-10; Z010, Zwick/Roell, Ulm, Germany). For the V40 taper, LFIT V40 heads (Stryker), for the 12/14 tapers, M-SPEC Metal heads (DePuy Synthes), both CoCr29Mo (\varnothing 36 mm) were used. The distal part of the models was fixed in a three-jaw lathe chuck (ZE-6353-125, Mack Werkzeuge, Sontheim Brenz, Germany) and inclined according to physiologically loading angles (10° adduction and 9° flexion; ISO 7206-4). Force-controlled cyclic loading (1 Hz) was applied via a PE piston (22 mm contact diameter) to the head using a servo-hydraulic testing machine (1 Hz; MiniBionix II, MTS, Eden Prairie, MN, US). A total of three activity levels were simulated that reflect situations of daily living [25]: $F_{L1}=230\text{--}2300$ N




Fig. 4. Setup for the digitalization of the prosthesis geometry using a 3D scanner.

Table 1

Geometrical properties for the 5 custom-made models ($n=3$). The reference cross sections of the neck geometry were taken 20 mm distally from the proximal taper surface (according to Fig. 1).

Specimen Type	Neck geometry			Male taper geometry			Head Type
	Cross section	a [mm]	b [mm]	Proximal \varnothing [mm]	Angle [°]	Length [mm]	
12/14	Round	12.98	–	12.550 ± 0.001	2.822 ± 0.002	11.2	M-Spec (DP)
V40	Round	13.05	–	11.300 ± 0.001	2.828 ± 0.001	10.8	LFIT V40 (ST)
L12/14	Round	12.43	–	12.550 ± 0.002	2.823 ± 0.003	13.7	M-Spec (DP)
12/14	Oval	12.60	11.50	12.550 ± 0.001	2.824 ± 0.002	11.2	M-Spec (DP)
12/14	Rounded square	13.23	9.30	12.550 ± 0.001	2.827 ± 0.003	11.2	M-Spec (DP)




Fig. 5. Calculation of the flexural rigidity for the taper and neck geometry. A 3D-scan of the prosthesis was obtained (a) and the corresponding point cloud derived (b). The precise geometry of the taper region was further gathered by a coordinate measuring machine. The flexural rigidity was calculated for each cross section of the prosthesis starting from the proximal plane.

(walking, 1 h), $F_{L2} = 230\text{--}4300\text{ N}$ (stair climbing, 0.5 h) and $F_{L3} = 230\text{--}5300\text{ N}$ (stumbling, 0.5 h).

The relative motion between the head and the stem component was measured contactless, utilizing six eddy-current sensors (Type ES05(78), MicroEpsilon, Ortenburg, Germany, resolution 50 nm). The sensors and corresponding reflectors (St37) were mounted into two holders that were rigidly fixed to the head and the models. The distance between sensors and their respective reflectors was continuously measured (Fig. 6a). Upon loading the corresponding relative motion was captured as a change in distance (measurement range 500 μm). Relative motion at the head center was computed using a coordinate transformation. Further details on the measurement setup are described elsewhere [39]. All measurements were performed within a temperature-controlled chamber containing four PID-controlled ceramic heat radiators (30 ± 0.1 °C;

Quantrol LC100, Jumo, Fulda, Germany). After testing, the pull-off forces of the heads were measured according to ISO 7206-10.

Testing of original Corail and Accolade prostheses was also performed to validate the TMZF FE-model of the Accolade proximal stem and taper as well as the manufactured model of the Corail proximal stem and taper. The distal parts of the stems were fixed at the resection plane using custom-made prosthesis-specific mounts (Fig. 6b), designed in accordance with ISO 7206-6.

2.4. Micromotion

When using non-destructive methods, it is only possible to determine the relative motion between the head and the stem, which is a combination of elastic deformation and micromotion. Micromotion can consequently be determined by subtracting the elastic




Fig. 6. Experimental setup used to measure relative motion between the head and stem. Six eddy current sensors are placed around the taper junction (a). Custom-designed holders were used to mount the prostheses in accordance with ISO standards (b).




Fig. 7. Notation for the directions of the measured relative motion at the head-stem junction with respect to a coordinate system fixed at the head center.

deformation from the measured relative motion. Relative motion was therefore measured for “near”-monoblocks of two of the models investigated (S12-14 round and V40 round). “Near”-monoblocks were produced by assembling respective heads and models with a very high axial assembly force of 70 kN (SCHENK, PC400M, Darmstadt, Germany). Higher forces could not be used since they caused plastic deformation of the models. Measured relative motion of the “near”-monoblocks consist nearly of the pure elastic deformation. After subtraction of the elastic deformation, pistoning, lateral and central rocking micromotion were combined to the total translational micromotion (Euclidean norm), while total rotational micromotion is described by the norm of screwing, central and lateral toggling (Fig. 7).

2.5. Finite element model

A total of 10 numerical models were created (ABAQUS 6.14-2, Dessault Systèmes, Vélizy-Villacoublay, France), 5 of those representing the modular head-stem taper junction with the individual stem-adapter designs (Fig. 3) and 5 to simulate corresponding monoblock versions analogue to the experimental study. In addition, the V40 models were simulated with two different material properties (Ti6Al4V: 120 GPa, TMZF: 85 GPa).

Each model consisted of the head and the proximal stem model as well as a simplified structure of the holders using beam elements (10 × 10 mm) for improved computational performances. The model was orientated analogue to the experimental setup fulfilling the ISO requirements. Sinusoidal load was applied over two cycles for each activity level due to the quasi-static simulation. Tangential behavior was used to discretize the interaction including finite sliding with a penalty friction formulation and a friction coefficient of $\mu=0.3$. All implant geometries were meshed with linear hexahedron elements (C3D8R). The contact area of the male taper (slave surface) was meshed with a seed size of 0.4 mm and of the female taper (master surface) with 0.8 mm. An increased mesh size was chosen for areas without surface contact. An analysis of convergence was performed to ensure a suitable mesh size. Monoblocks were modelled by rigidly combining the head and stem using a Boolean function. The ball head was translated on the male taper by the amount of secondary seating obtained from its corresponding modular model. Head (CoCr29Mo: 210 GPa) and stem material were reassigned to match their original geometries as before merging.

2.6. Statistical analysis

Pull-off force, total translational and rotational relative motion were compared statistically between activity level and stem or taper design, using a parametric analysis (one-way analysis of variance (ANOVA)) and post hoc test (Tukey), or a non-parametric analysis (Kruskal–Wallis) with the probability of a Type I error set

to $\alpha=0.05$ (SPSS Statistics 23; IBM Corp., Armonk, NY, US). All results are reported as mean and standard deviation (SD).

3. Results

3.1. Flexural rigidity of the taper and the neck region

The maximum taper rigidity for the 12/14 round taper was computed to be 188 Nm², whereas the long version of the 12/14 round taper had a slightly greater taper rigidity of 194 Nm². (Fig. 8a). The V40 round male taper showed a maximum computed taper rigidity of 126 Nm². All three taper designs showed the same distribution of flexural rigidity within the distal neck region. For the different stem geometries, taper rigidity was constant with a magnitude of 188 Nm² since the taper was similar for all geometries (Fig. 8b). The round proximal stem cross section showed similar flexural rigidities in both bending directions. An oval neck region led to more flexible properties in lateral direction compared to the central bending direction (−70%). Models with rounded square neck region proved to be most flexible in lateral direction when compared to the other models (−100%). The lower lateral flexural rigidity of the neck geometry was consistent with higher displacements observed in the finite element model (Fig. 9).

3.2. Total translational relative motion

Increasing activity levels caused a marked increase of total translational relative motion (Table 2; $p < 0.001$). Total translational relative motion of the 12/14 round model during walking was $16 \pm 0.6 \mu\text{m}$. For the smaller V40 round model, an increase to $19.2 \pm 0.9 \mu\text{m}$ was seen ($p < 0.001$; all activity levels). The increased taper length of the long L12/14 round model led to a significant increase of total translational relative motion compared to the reference taper ($22.9 \pm 0.4 \mu\text{m}$; $p < 0.001$; all activity levels).

The rounded square neck geometry showed an enlarged total translational relative motion compared to the round geometry ($17.3 \pm 1.1 \mu\text{m}$; $p < 0.001$). A similar difference was found for the oval neck design ($18.3 \pm 0.2 \mu\text{m}$; $p < 0.001$; all activity levels).

The results for the original Corail stem and the corresponding custom-made model (12/14 rounded square) showed slightly lower relative motion for the original stem during walking ($15.0 \mu\text{m}$ vs. $17.3 \mu\text{m}$). The numerical results matched the experimental results within 10% tolerance. For the activity level stumbling, a difference of only 0.05% was achieved (Table 2).

3.3. Total translational and rotational micromotion

Total translational and rotational micromotion in the junction (without elastic contribution of the neck) was higher for increasing activity levels (Fig. 10). For the 12/14 round reference model a total translational micromotion was computed within the range of 2.2–2.8 μm (walking-stumbling) and 0.003–0.010° for rotational motion. Similar amounts of total micromotion were computed for the 12/14 rounded square and 12/14 oval model (2.2–2.8 μm ; 0.004–0.010°). Prolongation of the taper (L12/14 round) led to increased total translational and rotational micromotion compared to the reference taper (2.3–3.0 μm and 0.005–0.012°). Models with V40 taper design exhibited increased total translational and rotational micromotion in the range of 2.9–4.5 μm (V40 round) and 3.3–5.2 μm (V40 round TMZF, Fig. 10a–b), respectively 0.002–0.016° and 0.006–0.022°.

3.4. Rocking and pistoning micromotion

By separating the total translational micromotion into rocking motion (radial to the taper axis) and pistoning motion (in axis of




Fig. 8. Flexural rigidity computed with a constant elastic modulus of 120 MPa for the different taper designs (a) and neck variations (b). Proximal stem models were clamped at 30 mm distance from the proximal taper plane.




Fig. 9. Finite element model of the proximal neck and taper region for the 12/14 rounded square (a) and 12/14 round neck geometry (b). The lower lateral flexural rigidity is highlighted by higher displacements in the neck.

Table 2 Experimentally and numerically determined relative motion of the different models.

Activity level [N]		Relative motion [μm]													
		12/14 round		L12/14 round		V40 round		12/14 oval		12/14 rounded square		12/14 round "near"-Monoblock		V40 round "near"-Monoblock	
		Exp.	Num.	Exp.	Num.	Exp.	Num.	Exp.	Num.	Exp.	Num.	Exp.	Num.	Exp.	Num.
FL1	230–2300	16.1 ± 0.6	17.4	22.9 ± 0.4	19.4	18.9 ± 0.9	19.9	18.3 ± 0.2	20.5	17.3 ± 1.1	18.8	14.7	16.5	17.2	17.7
FL2	230–4300	33.1 ± 0.8	33.7	46.3 ± 1.0	37.6	39.5 ± 1.1	38.2	37.4 ± 0.7	39.8	36.6 ± 0.4	36.7	30.0	32.5	35.9	34.9
FL3	230–5300	41.5 ± 0.8	41.7	58.1 ± 1.4	46.6	49.8 ± 1.3	47.1	47.2 ± 1.0	49.4	46.0 ± 0.3	45.5	37.8	40.5	44.8	43.5




Fig. 10. Total translational (a) and rotational (b) micromotion of the different models for the three activity levels.




Fig. 11. Rocking (a) and pistoning (b) micromotion of the different models for the three activity levels.

the taper) smaller taper diameters exhibited an increase in rocking motion of about 300% for all activity levels. Taper length or neck geometry did not show any differences in rocking micromotion (Fig. 11a).

Pistoning motion was observed in the range of $-2.2\mu\text{m}$ to $-2.8\mu\text{m}$ for the models using the 12/14 taper designs (12/14 round, rounded square, oval). Long tapers showed a slight increase in pistoning micromotion ($-2.3\mu\text{m}$ to $-3.0\mu\text{m}$). No effect of reduced taper diameter size (V40 round) on the amount of pistoning micromotion could be detected during walking. Pistoning micromotion increased by 15% compared to the reference model for higher activity levels. Lower elastic modulus resulted in an increase of pistoning micromotion of approximately 20% compared to the same taper design with the conventional material properties (Fig. 11b).

3.5. Taper strength analysis

The pull-off force for the models with the 12/14 taper and the long 12/14 taper design was similar ($1750 \pm 238\text{ N}$ re. $1679 \pm 97\text{ N}$; $p=1.00$). The V40 taper design showed a slightly elevated pull-off force ($2025 \pm 115\text{ N}$) compared to the 12/14 taper designs ($p=0.238$).

4. Discussion

Current research suggests that in-vivo material loss within modular taper junctions is attributable to mechanically assisted crevice corrosion (MACC; [40]). MACC can occur if two prerequisites are fulfilled: the presence of fluid in the body and micromotion between two mating surfaces [41]. The micromotion between the two surfaces can damage the passivation layer of the metal surface and leave the surface temporarily exposed to the surrounding body fluids. As a consequence, the prevention of micromotion is the key concept to minimize the impact of MACC [42].

Low-elastic modulus titanium composite stems (e.g. TMZF) were introduced to decrease stress shielding of the bone and improve femoral remodeling [43]. This beneficial characteristic, however, comes with drawbacks: the decreased elastic modulus may foster the expression of micromotion [36,39,44], which is confirmed in the present study. Another drawback of TMZF is the limited wear resistance under loading [45]. This is further supported by reports, showing less clinical adverse reactions for Ti6Al4V 11/13 tapers vs. TMZF V40 [30]. For bi-modular prostheses, TMZF

stems in combination with CoCr neck adapters were shown to exhibit extensive surface damage and CoCr material loss, again supporting that material is crucial for early implant failure [46,47].

The 30% lower elastic modulus of TMZF (TMZF with 85 GPa vs. Ti with 120 GPa) resulted in an increase of 15% in translational micromotion and of 30% in rotational micromotion (toggling). Separation of the translational micromotion showed that the influence of lower modulus of elasticity on micromotion was most prominent for pistoning motion (+30% vs. Ti6Al4V; Fig. 10) while rocking motion was not as extensively expressed (<5%). These findings corroborate with other publications reporting twice the amount of pistoning motion for TMZF compared to Ti at a loading amplitude of 3200 N (Accolade I: $12.5\mu\text{m}$ vs. Accolade II: $6.6\mu\text{m}$; [48]). The smaller micromotion values in the present study are due to the compensation of the elastic component from the relative motion by using numerical simulations.

Fretting corrosion within taper interfaces is a multifactorial problem that is affected not only by material choice, but also patient-specific parameters, handling by the surgeon and implant design [5,49]. The later one has been addressed in various reports, identifying the femoral stem taper flexural rigidity as a major factor for fretting corrosion [27–29,36]. In a retrieval analysis Arnholt and colleagues showed for total knee arthroplasties that the flexural rigidity of female tibial tapers was negatively correlated with fretting corrosion scores, while patient factors such as implantation time, patient weight and age showed no influence on corrosion [36]. Moreover, Higgs and colleagues showed in a retrieval study that stiffer tapers (range: $84.5\text{--}402.6\text{ Nm}^2$) exhibit less fretting damage, while the taper size (C taper vs. V40) had no effect on head fretting [29]. In contrast, the more flexible 11/13 taper sizes have been reported to show significantly higher taper fretting scores compared to 12/14 sizes even for a single manufacturer [30]. Consistent with these findings, the rocking micromotion was more than 300% higher for the V40 taper design compared to 12/14 sizes in the present study. The same trend can be observed for the toggling motion (+30% rotational micromotion [50]). Toggling can presumably lead to a cyclic opening and widening of the gap and subsequent fluid ingress [42] as well as severe plastic deformation and wear [51]. Increased corrosion has been found at the distal section of the taper region, possibly due to a combination of fluid ingress and enlarged bending moments [30], resulting in some cases even in catastrophic failure of the implant due to plastic deformation and abrasive wear of the taper surface [5,46].

Reduced taper length (mini vs. standard taper) was found to have a negative impact on fretting corrosion risk [52]. This finding is supported by the lower rocking motion observed for the long version of the 12/14 taper (−10%). The larger contact area in the long version might limit the bending loads via geometrical constraints and increased flexural rigidity. The higher pistoning motion for the long taper version might be explained by the reduced contact pressures as a consequence of the extended contact area. However, in this study, no differences in pull-off force were noticed that might be caused by lower contact pressures and higher contact areas for the long taper design compared to regular sized tapers. Fretting scores may vary to a great extent even for identical taper types due to variations in taper length, as they are adapted by manufacturers (e.g. range for 12/14 tapers: 10.8–14.2 mm; [7]). In the case of short tapers, the head-stem taper axes might be tilted due to higher bending moments resulting in an ongoing head rocking. In contrast, the increased pistoning of long tapers might be of secondary importance as no bending moments and widening of the gap are induced during translational motion. Whether a head-taper interface is more likely to be negatively affected by explicit head rocking or pistoning is still unclear. Retrieval studies show extensive damage pattern on opposite areas of the taper suggesting such a toggling effect [50]. In contrast, only minor differences between round, oval and rounded square designs regarding toggling motion were observed even for large variations in flexural rigidity of the neck designs. This could be different for heads with high offsets due to the higher bending and friction moments. Considering the current trend towards smaller and less stiff implants [10], the impact of bending moments might be even more crucial. On the other hand, the transition from round to rounded square neck geometries also offered advantages by reduction of the impingement risk, e.g. the original Corail neck vs. AMT Corail neck resulted in an increase of +10° in flexion and +18° in extension [12].

Micromotion increased significantly with higher activity level, indicating that patients with a high BMI are more affected by micromotion and possible implant failure [49]. Earlier studies indicate that fretting corrosion is directly linked to the amount of micromotion and contact pressure even if no critical value for micromotion can be derived [53]. The contact pressure within the taper junctions of the present study was initiated with 2000 N of assembly load for all specimen. Interestingly, in some cases, micromotion did not increase with higher activity levels (cyclic load of 5300 N), which might be explained by an additional seating of the head on the taper followed by less vulnerability for head toggling. If physiological friction moments are present, the prevalence for head toggling might even further increase and be expressed more clearly for taper junctions with lower contact pressures. In consideration of the fact that BMI and total number of performed THAs are increasing for most countries, the significance of micromotion and taper fretting could increase further.

The multifactorial nature of the problem is highlighted as contradicting reports have been published about the influence of single factors. Some studies did conclude that lower flexural rigidity was only found for specific material combinations [32,54] or could even lead to reduced damage scores [38]. These diverse findings [55] highlight the importance of isolated investigations of design, handling and patient-specific factors on the occurrence of MACC [8]. Retrieval studies typically address a mix of various factors such as geometry, surface morphology and material issues in combination with unknown parameters such as assembly and loading. The approach presented in this study using custom-made parts is in contrast well suited for the investigation of isolated parameters.

A limitation of the present study is the simplification to the proximal geometry of a prosthesis and minor deviations from the micro-grooved structure of the tapers used by the manufacturers.

Although geometrical parameters were kept identical to the original prostheses, manufacturing imperfections can have a significant impact on the quality of the taper junction [56]. This is also important to consider as manufacturing tolerances cannot be eliminated during an experimental study. Accordingly, those deviations cannot be considered in the numerical simulation as well as the precise material properties. Those values were obtained from literature, while variations in elastic modulus are known to exist for titanium alloys. Moreover, the simulations conducted in this study considered real-monoblock approaches. The experimentally tests were based on “near”-monoblocks as the crevice between the male and female taper depends on the taper angle difference.

5. Conclusion

In an attempt to increase impingement free range of motion, modern taper designs show a tendency to be thinner and shorter as traditional ones. This change reduces the rigidity of the modular taper junction [10], which has early been identified as a potential cause of fretting [27,33]. While the size, geometry and the material of the taper junction are not the only relevant factors contributing to risk of fretting corrosion [57], strong evidence has been presented that those factors are of high importance for the observed micromotion [5,8,42]. The influence of neck geometry as a risk factor was rather secondary in the present study. Favoring rounded square neck designs instead of smaller taper sizes could be an alternative to increase range of motion without enlarging the risk of fretting corrosion. The use of more elastic materials might have the similar effect as small tapers. Manufacturers should consider stiffer taper designs to avoid early implant failure.

Acknowledgments

The authors would like to thank DePuy Synthes for providing the prosthesis components.

Conflict of interests

All experimental studies were performed at the TUHH Hamburg University of Technology. The Institute of Biomechanics is receiving institutional support by DePuy Synthes but not related to this study. Prof. Morlock is a consultant to DePuy Synthes and serves on speaker bureaus for Aesculap, AORcon, Ceramtec, Corin, Lima, Mathys, Peter Brehm, DePuy Synthes, Zimmer-Biomet. The other authors have nothing to declare.

Ethical approval

Not required.

Funding

None.

References

- [1] Kluess D, Martin H, Mittelmeier W, Schmitz K-P, Bader R. Influence of femoral head size on impingement, dislocation and stress distribution in total hip replacement. *Med Eng Phys* 2007;29:465–71. doi:10.1016/j.medengphy.2006.07.001.
- [2] Matsushita A, Nakashima Y, Jingushi S, Yamamoto T, Kuraoka A, Iwamoto Y. Effects of the femoral offset and the head size on the safe range of motion in total hip arthroplasty. *J Arthroplasty* 2009;24:646–51. doi:10.1016/j.arth.2008.02.008.
- [3] Belvedere C, Cadossi M, Tamarri S, Lullini G, Ensini A, Cencioni A, et al. Impingement and range of motion in total hip replacement: a three-dimensional gait and video-fluoroscopic analysis. *Gait Posture* 2016;49:S19. doi:10.1016/j.gaitpost.2016.07.050.

- [4] Werner PH, Etema HB, Wit F, Morlock MM, Verheyen CCPM. Basic principles and uniform terminology for the head-neck junction in hip replacement. *HIP Int* 2015;25:115–19. doi:10.5301/hipint.5000204.
- [5] Morlock M, Bünthe D, Gührs J, Bishop N. Corrosion of the head-stem taper junction—are we on the verge of an epidemic?: Review article. *HSS J* 2017;13:42–9. doi:10.1007/s11420-016-9526-4.
- [6] Willmann G. Prinzip der konus-steckverbindung für keramische kugelpöfe bei huftendoprothesen. *Mater Wiss Und Werkst* 1993;319:315–19. doi:10.1002/mawe.19930240908.
- [7] Mueller U, Braun S, Schroeder S, Sonntag R, Kretzer JP. Same same but different? 12/14 stem and head tapers in total hip arthroplasty. *J Arthroplasty* 2017;32:3191–9. doi:10.1016/j.arth.2017.04.027.
- [8] Wight CM, Lanting B, Schemitsch EH. Evidence based recommendations for reducing head-neck taper connection fretting corrosion in hip replacement prostheses. *HIP Int* 2017;27:523–31. doi:10.5301/hipint.5000545.
- [9] Murray DW, Carr AJ, Bulstrode CJ. Which primary total hip replacement? *J Bone Joint Surg Br* 1995;77-B:520–7. doi:10.1302/0301-620X.77B4.7615593.
- [10] Porter DA, Urban RM, Jacobs JJ, Gilbert JL, Rodriguez JA, Cooper HJ. Modern trunnions are more flexible: a mechanical analysis of THA taper designs. *Clin Orthop Relat Res* 2014;472:3963–70.
- [11] Morlock MM, Bishop N, Huber G. Biomechanics of hip arthroplasty. In: Knahr K, editor. *Tribol. Total hip arthroplast.* Berlin, Heidelberg: Springer; 2011. p. 11–24. doi:10.1007/978-3-642-19429-0_2.
- [12] Vidalain J.-P., Selmi T.A.S., Beverland D, Young S, Board T, Boldt J, et al. The Corail® Hip system - a practical approach based on 25 years of experience. Berlin, Heidelberg: Springer; 2011.
- [13] Herrlin K, Selvik G, Pettersson H, Lidgren L. Range of motion caused by design of the total hip prosthesis. *Acta Radiol* 1988;29:701–4. doi:10.1177/028418518802900618.
- [14] Cuckler JM, Moore KD, Lombardi AV Jr, McPherson E, Emerson R. Large versus small femoral heads in metal-on-metal total hip arthroplasty. *J Arthroplasty* 2018;19:41–4. doi:10.1016/j.arth.2004.09.006.
- [15] Jacobs JJ, Urban RM, Gilbert JL, Skipor AK, Black J, Jasty M, et al. Local and distant products from modularity. *Clin Orthop Relat Res* 1995;319:94–105.
- [16] Urban RM, Jacobs JJ, Gilbert JL, Galante JO. Migration of corrosion products from modular hip prostheses. Particle microanalysis and histopathological findings. *J Bone Joint Surg Am* 1994;76:1345–59.
- [17] Gilbert JL, Buckley CA, Jacobs JJ, Bertin KC, Zernich MR. Intergranular corrosion-fatigue failure of cobalt-alloy femoral stems. A failure analysis of two implants. *JBS* 1994;76:110–15.
- [18] Cooper HJ, Della Valle CJ, Berger RA, Tetreault M, Paprosky WG, Sporer SM, et al. Corrosion at the head-neck taper as a cause for adverse local tissue reactions after total hip arthroplasty. *J Bone Jt Surg-Am* 2012;94:1655–61. doi:10.2106/JBJS.K.01352.
- [19] Kwon YM, Khormae S, Liow MHL, Tsai TY, Freiberg AA, Rubash HE. Asymptomatic pseudotumors in patients with taper corrosion of a dual-taper modular femoral stem MARS-MRI and metal ion study. *J Bone Jt Surg - Am* 2016;98:1735–40. doi:10.2106/JBJS.15.01325.
- [20] Gill IPS, Webb J, Sloan K, Beaver RJ. Corrosion at the neck-stem junction as a cause of metal ion release and pseudotumour formation. *J Bone Jt Surg - Br Vol* 2012;94-B:895–900. doi:10.1302/0301-620X.94B7.29122.
- [21] AOANJRR. Australian Orthopaedic Association National Joint Replacement Registry (AOANJRR). Hip, Knee & Shoulder Arthroplasty: 2017 Annual Report. Adelaide: 2017. Australian Orthopaedic Association.
- [22] Cook SD, Barrack RL, Baffes GC, Clemow AJ, Serekian P, Dong N, et al. Wear and corrosion of modular interfaces in total hip replacements. *Clin Orthop Relat Res* 1994;80–8. doi:10.1097/00003086-199401000-00013.
- [23] Silva M, Shepherd EF, Jackson WO, Dorey FJ, Schmalzried TP. Average patient walking activity approaches 2 million cycles per year. *J Arthroplasty* 2018;17:693–7. doi:10.1054/jarth.2002.32699.
- [24] Morlock M, Schneider E, Bluhm A, Vollmer M, Bergmann G, Müller V, et al. Duration and frequency of every day activities in total hip patients. *J Biomech* 2001;34:873–81. doi:10.1016/S0021-9290(01)00035-5.
- [25] Bergmann G, Graichen F, Rohlmann A, Bender A, Heinlein B, Duda GN, et al. Realistic loads for testing hip implants. *Biomed Mater Eng* 2010;20:65–75. doi:10.3233/BME-2010-0616.
- [26] Gilbert JL, Buckley CA, Jacobs JJ. In vivo corrosion of modular hip prosthesis components in mixed and similar metal combinations. The effect of crevice, stress, motion, and alloy coupling. *J Biomed Mater Res* 1993;27:1533–44. doi:10.1002/jbm.820271210.
- [27] Goldberg JR, Gilbert JL, Jacobs JJ, Bauer TW, Paprosky W, Leurgans S. A multicenter retrieval study of the taper interfaces of modular hip prostheses. *Clin Orthop Relat Res* 2002;401:149–61.
- [28] Kao Y-Y, Koch CN, Wright TM, Padgett DE. Flexural rigidity, taper angle, and contact length affect fretting of the femoral stem trunnion in total hip arthroplasty. *J Arthroplasty* 2016;1–5. doi:10.1016/j.arth.2016.02.079.
- [29] Higgs GB, MacDonald DW, Gilbert JL, Rinnac CM, Kurtz SM, Chen AF, et al. Does taper size have an effect on taper damage in retrieved metal-on-polyethylene total hip devices? *J Arthroplasty* 2016;31:277–81. doi:10.1016/j.arth.2016.06.053.
- [30] Tan SC, Teeter MG, Del Balso C, Howard JL, Lanting BA. Effect of taper design on trunnionosis in metal on polyethylene total hip arthroplasty. *J Arthroplasty* 2015;30:1269–72. doi:10.1016/j.arth.2015.02.031.
- [31] Dyrkacz RMR, Brandt JM, Morrison JB, O'Brien ST, Ojo OA, Turgeon TR, et al. Finite element analysis of the head-neck taper interface of modular hip prostheses. *Tribol Int* 2014;91:1–8. doi:10.1016/j.triboint.2015.01.016.
- [32] Kurtz SM, Kocagöz SB, Hanzlik Ja, Underwood RJ, Gilbert JL, Macdonald DW, et al. Do ceramic femoral heads reduce taper fretting corrosion in hip arthroplasty? A retrieval study. *Clin Orthop Relat Res* 2013;471:3270–82. doi:10.1007/s11999-013-3096-2.
- [33] Jennings JM, Dennis DA, Yang CC. Corrosion of the head-neck junction after total hip arthroplasty. *J Am Acad Orthop Surg* 2016;24:349–56. doi:10.5435/JAAOS-D-15-00111.
- [34] Nassif N, Nawabi DH, Stoner K, Elpers M, Wright T, Padgett DE. Taper design affects failure of large-head metal-on-metal total hip replacements. *Clin Orthop Relat Res* 2014;472:564–71. doi:10.1007/s11999-013-3115-3.
- [35] Triantafyllopoulos GK, Elpers ME, Burket JC, Esposito CI, Padgett DE, Wright TM. Otto Aufranc award: large heads do not increase damage at the head-neck taper of metal-on-polyethylene total hip arthroplasties. *Clin Orthop Relat Res* 2015 Epub ahead of print. doi:10.1007/s11999-015-4468-6.
- [36] Arnholt CM, MacDonald DW, Tohfafarosh M, Gilbert JL, Rinnac CM, Kurtz SM, et al. Mechanically assisted taper corrosion in modular TKA. *J Arthroplasty* 2014;29:205–8. doi:10.1016/j.arth.2013.12.034.
- [37] Arnholt CM, MacDonald DW, Underwood RJ, Guyer EP, Rinnac CM, Kurtz SM, et al. Do stem taper microgrooves influence taper corrosion in total hip arthroplasty? A matched cohort retrieval study. *J Arthroplasty* 2016. doi:10.1016/j.arth.2016.11.018.
- [38] Pourzal R, Bs DJH, Ba NQH, Urban RM, Levine BR, Jacobs JJ, et al. Does surface topography play a role in taper damage in head-neck modular junctions? *Clin Orthop Relat Res* 2016. doi:10.1007/s11999-016-4933-x.
- [39] Haschke H, Jauch-Matt SY, Sellenschloh K, Huber G, Morlock MM. Assembly force and taper angle difference influence the relative motion at the stem-neck interface of bi-modular hip prostheses. *Proc Inst Mech Eng Part H J Eng Med* 2016;230:690–9. doi:10.1177/0954411916648717.
- [40] Mali SA, Gilbert JL. Correlating fretting corrosion and micromotions in modular tapers: test method development and assessment. Modularity and tapers in total joint replacement devices, 100 Barr Harbor Drive, PO Box C700, 19428-2959. West Conshohocken, PA: ASTM International; 2015. p. 259–82. doi:10.1520/STP159120140136.
- [41] Gilbert JL, Mali S, Urban RM, Silverton CD, Jacobs JJ. In vivo oxide-induced stress corrosion cracking of Ti-6Al-4 V in a neck-stem modular taper: emergent behavior in a new mechanism of in vivo corrosion. *J Biomed Mater Res - Part B Appl Biomater* 2012;100 B:584–94. doi:10.1002/jbm.b.31943.
- [42] Pourzal R, Lundberg HJ, Hall DJ, Jacobs JJ. What factors drive taper corrosion? *J Arthroplasty* 2018;33:2707–11. doi:10.1016/j.arth.2018.03.055.
- [43] Head WC, Bauk DJ, Emerson RH. Titanium as the material of choice for cementless femoral components in total hip arthroplasty. *Clin Orthop Relat Res* 1995;311:85–90.
- [44] Jauch SY, Huber G, Haschke H, Sellenschloh K, Morlock MM. Design parameters and the material coupling are decisive for the micromotion magnitude at the stem-neck interface of bi-modular hip implants. *Med Eng Phys* 2014;36:300–7. doi:10.1016/j.medengphy.2013.11.009.
- [45] Yang X, Hutchinson CR. Corrosion-wear of β -Ti alloy TMZF (Ti-12Mo-6Zr-2Fe) in simulated body fluid. *Acta Biomater* 2016;42:429–39. doi:10.1016/j.actbio.2016.07.008.
- [46] Buentje D, Huber G, Bishop N, Morlock M. Quantification of material loss from the neck piece taper junctions of a bimodular primary hip prosthesis. A retrieval study from 27 failed Rejuvenate bimodular hip arthroplasties. *Bone Jt J* 2015;97B:1350–7. doi:10.1302/0301-620X.97B10.35342.
- [47] Su SL, Koch CN, Nguyen TM, Burket JC, Wright TM, Westrich GH. Retrieval analysis of neck-stem coupling in modular hip prostheses. *J Arthroplasty* 2017;32:2301–6. doi:10.1016/j.arth.2017.02.016.
- [48] Swaminathan V, Scholl L, Lee R, Faizan A, Thakore M, TenHuisen K, et al. Simultaneous hip head-stem taper junction measurements of electrochemical corrosion and micromotion: a comparison of taper geometry and stem material. Modularity and tapers in total joint replacement devices, 100 Barr Harbor Drive, PO box C700. West Conshohocken, PA: ASTM International; 2015. 19428-2959 p. 321–35. doi:10.1520/STP159120140150.
- [49] Morlock MM. The taper disaster - How could it happen? *HIP Int* 2015;25:339–46. doi:10.5301/hipint.5000269.
- [50] Esposito CI, Wright TM, Goodman SB, Berry DJ. What is the trouble with trunnions? *Clin Orthop Relat Res* 2014;472:3652–8. doi:10.1007/s11999-014-3746-z.
- [51] Bishop N, Witt F, Pourzal R, Fischer A, Rüttschi M, Michel M, et al. Wear patterns of taper connections in retrieved large diameter metal-on-metal bearings. *J Orthop Res* 2013;31:1116–22. doi:10.1002/jor.22326.
- [52] Panagiotidou A, Meswania J, Hua J, Muirhead-Allwood S, Hart A, Blunn G. Enhanced wear and corrosion in modular tapers in total hip replacement is associated with the contact area and surface topography. *J Orthop Res* 2013;31:2032–9. doi:10.1002/jor.22461.
- [53] Baxmann M, Jauch SY, Schilling C, Blömer W, Grupp TM, Morlock MM. The influence of contact conditions and micromotions on the fretting behavior of modular titanium alloy taper connections. *Med Eng Phys* 2013;35:676–83. doi:10.1016/j.medengphy.2012.07.013.
- [54] Lundberg HJ, Ha NQ, Hall DJ, Urban RM, Levine BR, Pourzal R. Contact mechanics and plastic deformation at the local surface topography level after assembly of modular head-neck junctions in modern total hip replacement devices. Modularity and tapers in total joint replacement devices, 100 Barr Harbor Drive, PO box C700. West Conshohocken, PA: ASTM International; 2015. 19428-2959 p. 59–82. doi:10.1520/STP159120140148.

- [55] Baleani M, Erani P, Bordini B, Zuccheri F, Makosa MK, De Pasquale D, et al. In vivo damage of the head-neck junction in hard-on-hard total hip replacements: effect of femoral head size, metal combination, and 12/14 taper design. *Materials (Basel)* 2017;10. doi:[10.3390/ma10070733](https://doi.org/10.3390/ma10070733).
- [56] Shareef N, Levine D. Effect of manufacturing tolerances on the micromotion at the Morse taper interface in modular hip implants using the finite element technique. *Biomaterials* 1996;17:623–30. doi:[10.1016/0142-9612\(96\)88713-8](https://doi.org/10.1016/0142-9612(96)88713-8).
- [57] Rowan FE, Wach A, Wright TM, Padgett DE. The onset of fretting at the head-stem connection in hip arthroplasty is affected by head material and trunnion design under simulated corrosion conditions. *J Orthop Res* 2018;36:1630–6. doi:[10.1002/jor.23813](https://doi.org/10.1002/jor.23813).