

**Materials and methods.** According to European and Bulgarian legislation each X-ray has to be equipped with enough protection clothing. Every year these clothing have to be examined for defects which could lead to over exposure of the staff. A simple method for testing and rejection of the clothing was presented from Brian Philips et al. This method comprises several steps for examination together with rejection criteria for lead apron, thyroid shield and Gauntlet.

**Results.** Because of the daily use in most of the clothes defects were observed. The size of some of the defects could cause overexposure and overrun of the year dose limit of 20 mSv. Therefore these clothes were rejected and replaced with new.

**Conclusions.** Examination of the lead clothing must be included in every quality assurance program. This will keep the occupational exposure below the year effective dose limit.

<https://doi.org/10.1016/j.ejmp.2018.12.018>

### Automated dose tracking in CT of the chest

E. Georgiev<sup>a</sup>, I. Dyakov<sup>b</sup>

<sup>a</sup>Acibadem City Clinic Tokuda Hospital, Bulgaria

<sup>b</sup>Acibadem City Clinic Cancer Centre, Bulgaria

E-mail: [emil.georgiev@acibademcityclinic.bg](mailto:emil.georgiev@acibademcityclinic.bg)

**Purpose.** To present a detailed optimization process of automated radiation dose tracking data in standard CT chest examination for multi detector computed tomography (MDCT). Tasks resulting from such a goal are related with correct processing of collected data, deep analysis and interpretation of the data, identification of incorrect clinical practice and to compare the results with another radiology department practice.

**Materials and methods.** A single retrospective study was conducted involving regular enrolled patients for routine CT chest examination in Acibadem City Clinic Tokuda Hospital. All MDCT exams were performed on a 64-detector row computed tomography (LightSpeed VCT, GEHC) without any iterative reconstructions. The tube voltage (kV), slice thickness (mm), rotation time (s/rot) and table speed (mm/s) were periodically altered parameters according to radiation dose reduction without compromising the quality of the image. In the course of the study a web-based dose management software was used (DoseWatch, GEHC) which allows all MDCT chest examinations to be properly tracked and radiation doses to be collected.

**Results.** After the step by step optimization the CTDI<sub>vol</sub> and DLP were decreased to 9,20 mGy and 373,77 mGy.cm respectively while maintaining good diagnostic image quality. In comparison with Acibadem City Clinic Cancer Centre (Optima CT660, SS40 ASIR, GEHC) these values are slightly higher which is explained by the fact that in Optima CT660 an iterative reconstruction algorithm was systematically used (CTDI<sub>vol</sub>6,43 mGy and DLP 265,2 mGy.cm).

**Conclusions.** The CT chest radiation dose was reduced with 56% (CTDI<sub>vol</sub>) and 63% (DLP) with preserving the diagnostic image quality. A significant CT chest dose reduction was achieved even without using any iterative reconstruction algorithm.

<https://doi.org/10.1016/j.ejmp.2018.12.019>

### An approach of modelling of irregular masses

Galya Gospodinova, Kristina Bliznakova

Laboratory of Computer Simulations in Medicine, Technical University of Varna, Bulgaria

E-mail: [kristina.bliznakova@tu-varna.bg](mailto:kristina.bliznakova@tu-varna.bg)

The goal of this study is to create and evaluate a methodology for generation of realistic 3D computational models of breast tumours with irregular shapes.

**Methods.** The methodology for the creation of breast masses consists of two steps: (a) creation of the initial diffusive tumour shape, by choosing Brownian motion or nearest neighbour random walk algorithm, (b) creation of a solid tumour shape by applying a set of 3D filters, and morphological operations. The initial models were smoothed by applying a set of image processing methods. Thereafter, projection images of these 3D lesions were generated by using an in-house developed software, capable to simulate the X-ray transport through the computational phantoms. Subsequently, a technique for embedding the simulated masses in patient mammography images was developed and applied.

**Results.** Thirty irregular masses with different sizes and shapes were generated, and projection images were simulated. The realism of the projected masses on patient images was evaluated by comparison of extracted features, such as the exponential parameter of the power spectrum, fractal dimension and other statistical parameters to these extracted from patient data. The obtained results confirmed that the methodology is capable of producing realistic 3D tumour models with user defined sizes and irregular shapes.

**Conclusions.** The methodology will be used to generate unique and realistic in shape and size computational models of breast adenoma, intraductal papilloma, cysts and duct hyperplasia. These computational models are stored in an open source database to be used by all professionals working toward the creation of new technologies for breast-screening and diagnosing.

**Acknowledgements.** This research is supported by the Bulgarian National Science Fund under grant agreement DN17/2. This project also has received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement No 692097.

<https://doi.org/10.1016/j.ejmp.2018.12.020>

### Breast awareness mobile application

Irena Georgieva

Laboratory of Computer Simulations in Medicine, Technical University of Varna, Bulgaria

E-mail: [irenayordgeorgieva@gmail.com](mailto:irenayordgeorgieva@gmail.com)

The main purpose of this work is to present the development of a mobile application to assist the women with breast cancer prevention. The mobile application is implemented using React Native - a JavaScript framework for building natively rendering mobile applications for iOS and Android. The RESTful API is implemented using Node.js - an open-source, asynchronous, event driven, cross-platform JavaScript run-time environment. MongoDB - a NoSQL, document-oriented database is used as a database. The platform starts with a questionnaire related to the user and her family. After the quiz, the application assesses the user's risk of breast cancer and provides advices how often a specialist should be visited. The questions, as well as all other information supplied by the application, are currently under review and evaluation by radiologists. Under development is as well a system for archiving the user's medical examination results. In addition, all the necessary information about prophylaxis, self-exam, symptoms, stages, risk factors, etc. is provided in an interactive way. For instance, there is a dedicated panel for implementing a breast self-exam, which is important action towards finding first signs of any breast abnormality. The provided drawings and running text assist the users to easily perform this